

ReAlign 3003

| | | | |
|------------------|-------------|-----------------|---|
| Batch ID: | TAH15003003 | Test ID: | T000005947 |
| Reported: | 16-Sep-2019 | Method: | Arsenic = Arsenic EPA 6020A (mod), Cadmium = Cadmium EPA 6020A (mod), Lead = Lead EPA 6020A (mod), Mercury = Mercury EPA 6020A (mod) |
| Type: | Other | | |
| Test: | Metals | | |

HEAVY METALS

| Compound | Reporting Limit (ppm) | Result (ppm) |
|----------|-----------------------|--------------|
| Arsenic | 0.01 | 0.01 |
| Cadmium | 0.01 | <0.01 |
| Lead | 0.01 | <0.01 |
| Mercury | 0.01 | <0.01 |

FINAL APPROVAL

Sam Smith
16-Sep-2019
4:01 PM

PREPARED BY / DATE

Chris Jungling
16-Sep-2019
4:11 PM

APPROVED BY / DATE

Testing results are based solely upon the sample submitted to Botanacor Laboratories, LLC, in the condition it was received. Botanacor Laboratories, LLC warrants that all analytical work is conducted professionally in accordance with all applicable standard laboratory practices using validated methods. Data was generated using an unbroken chain of comparison to NIST traceable Reference Standards and Certified Reference Materials. This report may not be reproduced, except in full, without the written approval of Botanacor Laboratories, LLC.